

**Certificate of Analysis**

**Product:** Dichloromethane  
**Grade:** HPLC  
**Part No.'s:** 1913-5

<b>Lot:</b>	<b>I3R110</b>
<b>Mfg. Date:</b>	<b>2018-03-27</b>
<b>Exp. Date:</b>	<b>3 years after Mfg. Date</b>

*Meets ACS Specification*  
**Ultraviolet Spectrophotometry**

**Maximum UV Absorbance:**

Wavelength	Unit	Specification (Max)	Results
233	nm	1.00	0.35
235	nm	0.50	0.22
240	nm	0.15	0.07
254	nm	0.01	0.002
280		0.01	0.001
UV Cutoff	nm	233	<233

**LC Gradient Suitability**

Test	Unit	Specification	Results
Assay (by GC)	%	Min. 99.9	99.9
Color (APHA)		Max. 10	3
Water	%	Max. 0.02	0.003
Residue After Evaporation	ppm	Max. 2.0	<1
Titration acid	mEq/g	Max. 0.0003	0.00003
Free Halogens		To Pass Test	Pass

Contains Stabilizer (Amylene) 15~200ppm

Created By:



Colette Batog  
 Technical Support