

Certificate of Analysis

Product: Methanol
Grade: LC-MS
Part No: 1935-5

Lot:	L89201
Mfg. Date:	2021-08-09
Exp. Date:	3 years after Mfg. Date

Manufactured by Duksan Pure Chemicals

LC-MS Suitability

Test	Unit	Specification	Results
ESI+ mode (as Reserpine)	ppb	Max. 50	<50
ESI- mode (as 4-Nitrophenol)	ppb	Max. 50	<50
Assay (by GC)	%	Min. 99.9	99.9
Color (APHA)		Max. 10	3
Water	%	Max. 0.03	0.007
Residue After Evaporation	ppm	Max. 1	<1
Titration Acid	mEq/g	Max. 0.0003	0.0001
Titration Base	mEq/g	Max. 0.0002	0.00003
Ultraviolet Spectrophotometry			
Wavelength	Unit	Maximum UV Absorbance	
220	nm	0.25	0.15
230	nm	0.15	0.07
254	nm	0.02	0.01
280	nm	0.01	0.002
LC Gradient Suitability		To Pass Test	Pass

Metal Impurities

Metal	Unit	Specification (Max)	Results
Aluminum (Al)	ppb	25	<1
Barium (Ba)	ppb	5	<1
Cadmium (Cd)	ppb	5	<1
Calcium (Ca)	ppb	25	<1
Chromium (Cr)	ppb	5	<1
Cobalt (Co)	ppb	5	<1
Copper (Cu)	ppb	5	<1

Iron (Fe)	ppb	5	<1
Magnesium (Mg)	ppb	25	<1
Manganese (Mn)	ppb	5	<1
Lithium (Li)	ppb	5	<1
Nickel (Ni)	ppb	5	<1
Potassium (K)	ppb	25	<1
Silver (Ag)	ppb	5	<1
Sodium (Na)*	ppb	50	<1
Tin (Sn)	ppb	5	<1

* May change over time

Created By:
Fernando Fana
Quality Associate