

**Certificate of Analysis**

**Product:** Water  
**Grade:** LC-MS  
**Part No.'s:** 1939-5  
**CAS No:** 7732-18-5

<b>Lot:</b>	<b>N2K110</b>
<b>Mfg. Date:</b>	<b>2023-02-20</b>
<b>Exp. Date:</b>	<b>2 years after Mfg. Date</b>

Manufactured by Duksan Pure Chemicals

**LC-MS Suitability**

Test	Unit	Specification (Max)	Results
ESI+ mode (as Reserpine)	ppb	50	<50
ESI- mode (as 4-Nitrophenol)	ppb	50	<50
Residue After Evaporation	ppm	1	<1
Ultraviolet Spectrophotometry			
Wavelength	Unit	Maximum UV Absorbance	
200	nm	0.01	<0.01
250 - 400	nm	0.005	<0.005
Gradient Elution Test			
Test	Unit	Specification (Max)	
Largest Eluted Peak(254nm)	mAU	1	<1
Largest Eluted Peak(210nm)	mAU	5	<5
Fluorescence Trace Impurities			
Test	Unit	Specification (Max)	
At 450nm Emissions	ppb	0.1	<0.1
At Emissions Maximum for Impurities	ppb	0.2	<0.2

**Metal Impurities**

Metal	Unit	Specification (Max)	Results
Aluminum (Al)	ppb	25	<1
Barium (Ba)	ppb	5	<1
Cadmium (Cd)	ppb	5	<1
Calcium (Ca)	ppb	25	<1
Chromium (Cr)	ppb	5	<1

Cobalt (Co)	ppb	5	<1
Copper (Cu)	ppb	5	<1
Iron (Fe)	ppb	5	<1
Lead (Pb)	ppb	5	<1
Magnesium (Mg)	ppb	25	<1
Manganese (Mn)	ppb	5	<1
Lithium (Li)	ppb	5	<1
Nickel (Ni)	ppb	5	<1
Potassium (K)	ppb	25	<1
Silver (Ag)	ppb	5	<1
Sodium (Na)*	ppb	50	<6
Tin (Sn)	ppb	5	<1
Zinc (Zn)	ppb	5	<1

\* May change over time

Tested to ACS specifications

Filtered by 0.2um filter

Created By:

James Osborn  
Quality Supervisor