

**Acetonitrile**  
**FOR GRADIENT**

[75-05-8] ( CH<sub>3</sub>CN )FW:41.05

**Product code : 1910-5**

**HPLC Grade**

**Lot-O2L201**

TESTS	UNIT	SPECIFICATION	RESULTS
UV Absorbance 195 nm		Max. 0.15	0.07
UV Absorbance 200 nm		Max. 0.05	0.03
UV Absorbance 205 nm		Max. 0.04	0.02
UV Absorbance 210 nm		Max. 0.02	0.017
UV Absorbance 220 nm		Max. 0.01	0.008
UV Absorbance 254 nm		Max. 0.005	0.001
UV Cutoff	nm	Max. 190	<190
Gradient Elution test			
- Largest eluted peak(254nm)	mAU	Max. 0.5	<0.5
- Largest eluted peak(210nm)	mAU	Max. 2	<2
Fluorescence Trace Impurities			
- at 450nm Emission	ppb	Max. 0.3	<0.3
- at Emission Maximum for impurities	ppb	Max. 0.5	<0.5
Assay (by GC)	%	Min. 99.9	99.9
Color (APHA)		Max. 10	2
Water	%	Max. 0.01	0.001
Residue after evaporation	ppm	Max. 1	<1
Titrateable Acid	mEq/g	Max. 0.0008	0.0006
Titrateable Base	mEq/g	Max. 0.0006	0.00001
Filtered by 0.2um filter			

Mfg. Date : 2024-02-21

Exp. Date : 3 years after Mfg. Date

Test Method : ACS 11th Edition

Tested By : Jong-sun,Won

**DUKSAN PURE CHEMICALS CO., LTD**

Ansan | 53, Sinwon-ro 133beon-gil, Danwon-gu, Ansan-si, Gyeonggi-do, Korea  
 jinCheon | 13-37, Sansusandan 1-ro, Deoksan, Jincheon-gun, Chungcheongbuk-do, Korea  
 TEL : +82-31-495-6886 FAX : +82-31-495-5335  
 World Wide Web : WWW.DUKSAN.KR  
 E-mail : QC@DUKSAN.KR



Quality Assurance Director